



Providing Choice & Value
Generic CT and MRI Contrast Agents



CONTACT REP

AJNR

Reply:

D.F. Kallmes, H.J. Cloft, J.B. White and C.G.M. Ken

AJNR Am J Neuroradiol 2010, 31 (4) E42

doi: <https://doi.org/10.3174/ajnr.A2030>

<http://www.ajnr.org/content/31/4/E42>

This information is current as
of July 7, 2025.

Reply:

We greatly appreciate Dr Jou's interest in our work. His approach is innovative. We are gratified to see that the major determinant of coil softness, the diameter of the primary coil structure, is the same in his analysis and ours. We concur that empiric testing of coil softness remains problematic, but both his calculations and ours appear in line with our own clinical experience.

D.F. Kallmes

H.J. Cloft

Mayo Clinic

Rochester, Minnesota

J.B. White

Texas A&M Health Science Center

College Station, Texas

C.G.M. Ken

San Mateo, California

DOI 10.3174/ajnr.A2030