


On-line Table: Comparison of MRI techniques with DSA for aneurysm measurements

	Mean (mm)	SD	CV (100%)	Bias	LOA	ICC
ICA (<i>n</i> = 32)						
TOF-MRA	10.49	1.57	14.73	−0.19	(−3.26–2.89)	0.97 (0.96–0.98)
CE-MRA	10.96	0.66	6.12	0.26	(−1.03–1.55)	0.98 (0.98–0.99)
3D-BB MRI	10.67	0.60	5.61	0.03	(−1.15–1.21)	0.99 (0.99–0.99)
MCA (<i>n</i> = 12)						
TOF-MRA	7.44	1.44	18.34	0.43	(−2.39–3.25)	0.88 (0.77–0.94)
CE-MRA	7.97	0.66	8.43	−0.10	(−1.39–1.19)	0.98 (0.96–0.99)
3D-BB MRI	7.84	0.55	7.03	0.04	(−1.04–1.12)	0.98 (0.97–0.99)
AcomA (<i>n</i> = 10)						
TOF-MRA	7.31	1.31	17.57	0.18	(−2.39–2.75)	0.94 (0.87–0.97)
CE-MRA	7.80	0.68	9.04	−0.30	(−1.63–1.03)	0.98 (0.96–0.99)
3D-BB MRI	7.73	0.49	6.55	−0.23	(−1.19–0.73)	0.99 (0.97–0.99)
PcomA and PCA (<i>n</i> = 8)						
TOF-MRA	7.92	1.21	15.01	0.15	(−1.61–2.52)	0.90 (0.79–0.96)
CE-MRA	8.58	0.50	6.21	−0.51	(−1.49–0.47)	0.97 (0.78–0.99)
3D-BB MRI	8.09	0.51	6.33	−0.02	(−1.02–0.98)	0.98 (0.96–0.99)
VBA (<i>n</i> = 6)						
TOF-MRA	6.73	0.88	13.16	−0.02	(−1.74–1.70)	0.89 (0.73–0.96)
CE-MRA	6.91	0.55	8.23	−0.19	(−1.27–0.89)	0.95 (0.88–0.98)
3D-BB MRI	6.87	0.51	7.56	−0.15	(−1.15–0.85)	0.96 (0.90–0.98)

Note:—AcomA indicates anterior communicating artery; PcomA, posterior communicating artery; VBA, vertebrobasilar artery; PCA, posterior cerebral artery.



ON-LINE FIGURE. Bland-Altman plots of aneurysm height (A–C), width (D–F), and neck (G–I) measurements of TOF-MRA (A, D, and G), CE-MRA (B, E, and H), and 3D-BB MR (C, F, and I) versus 3DRA as a reference standard. Solid horizontal lines define the reference standard, and upper and lower dashed lines define LOA.